

ADS-300

ANESTHESIA WORKSTATION WITH AVS-180 Ventilator + 2 Gas



PARAMETERS	AVS-180
Mode of Operation	CMV, ACMV, SIMV, PLV, BAG & Standby.
Patient Type	Adult and Child [Selectable]
I: E Ratio	4:1 to 1:4
Plateau	0 to 50%
Trigger	0.5 to 5.0 cmH ₂ O
Tidal Volume	50 to 1500 ml.
Respiratory Rate	1 to 40 Bpm
Ti	0.3 to 8.0 Sec
Pi	5 to 60 cmH ₂ O
PEEP	-
Pressure Support	-
Ti _{max}	-
Backup RR	-
Backup Mode	CMV
Monitoring Parameters	Tidal Volume, Minute Volume, RR, FIO ₂ , Pi, P Mean, P Plat, PEEP Obtained.
Alarms	Pressure Low and High, RR Low and High, FIO ₂ Low and High, Active patient, Patient disconnected, Oxygen Failure Low battery indication, Minute Volume low-High.
Loop Types	Flow vs Pressure, Flow vs Volume & Volume vs Pressure [Selectable Any Two]
Graph Types	Flow, Pressure & Volume [Selectable Any Two]
Fresh Gas Decoupling	Enabled
Flow Sensor	External
Display	7" TFT with touch Screen.
Power Input	20 W
Driving Gas	Oxygen
Gas Input Pressure	3-6 Bar (Pipeline) / 4-150 Bar (Cylinder)
Battery Backup	2 hours (Lithium-ion)
Flow Meter Type	2 Gas (4 Tubes)



Manufactured by: **Alveonic Medical Systems P Ltd,**
Coimbatore – 641008 India.

Ph: +91 - 9363993324 / +91 - 9597399973

Email: info@alveonic.com | Website: www.alveonic.com